

Updated calibration model for ALPHA II

From today 23/10/2025 this is the only valid method to use, and all further analysis shall be done using this new model, until anything is announced.

(FIS_IBU Fluor_finder vs 251007.q2 file)